

To: Arguto, William[Arguto.William@epa.gov]
From: binetti, victoria
Sent: Mon 1/13/2014 8:40:01 PM
Subject: Fw: ICLN Emergency Method Validation

From: Allgeier, Steve
Sent: Monday, January 13, 2014 3:33:07 PM
To: binetti, victoria
Cc: Hedrick, Elizabeth
Subject: ICLN Emergency Method Validation

In addition to some data, the following info would need to be provided. Maybe check in with Eastman or the State to see if this documentation exists?

- 5.1.4 A written procedure which includes all the following elements must be prepared:
- Intended use and criteria for use
 - Assay principle and safety precautions,
 - Acceptable sample types, sample collection method, preparation, preservation, storage and transportation conditions ,
 - Description of reagents supplied or directions for preparation and QC,
 - Description of quality controls to be used
 - Detailed instructions on how to perform the method,
 - Detailed instructions on how to interpret and report the result,
 - Any limitations of the method that are known or suspected,
 - A summary of the performance characteristics of the method,
 - A statement on the target uncertainty of measurements for each analyte (analytical goal for accuracy) in order for the values to be fit for their intended purpose.

Thanks,

Steve

Steven C. Allgeier

USEPA, OW, OGWDW

Water Security Division

26 W. Martin Luther King Dr.

Cincinnati, OH 45268

Allgeier.Steve@EPA.GOV

Phone: 513-569-7131

Fax: 513-569-7191